
Number 3

Volume 10



March 2019

Indian Journal of Public Health Research & Development

An International Journal

SCOPUS IJPHRD CITATION SCORE

Indian Journal of Public Health Research and Development Scopus coverage years: from 2010 to 2018 Publisher: R.K. Sharma, Institute of Medico-Legal Publications ISSN:0976-0245E-ISSN: 0976-5506 Subject area: Medicine: Public Health, Environmental and Occupational Health Cite Score 2017- 0.03

Cite Score 2017- 0.03 SJR 2017- 0.108 SNIP 2017- 0.047



Website: www.ijphrd.com

Indian Journal of Public Health Research & Development

EXECUTIVE EDITOR

Prof Vidya Surwade

Associate Professor, Dr Baba Saheb Ambedkar, Medical College & Hospital, Rohinee, Delhi

INTERNATIONAL EDITORIAL ADVISORY BOARD

- Dr. Abdul Rashid Khan B. Md Jagar Din, (Associate Professor)
 Department of Public Health Medicine, Penang Medical College, Penang, Malaysia
- Dr. V Kumar (Consulting Physician)
 Mount View Hospital, Las Vegas, USA
- 3. Basheer A. Al-Sum,

Botany and Microbiology Deptt, College of Science, King Saud University, Riyadh, Saudi Arabia

- Dr. Ch Vijay Kumar (Associate Professor)
 Public Health and Community Medicine, University of Buraimi, Oman
- Dr. VMC Ramaswamy (Senior Lecturer)
 Department of Pathology, International Medical University, Bukit Jalil, Kuala Lumpur
- Kartavya J. Vyas (Clinical Researcher)
 Department of Deployment Health Research,
 Naval Health Research Center, San Diego, CA (USA)
- Prof. PK Pokharel (Community Medicine)
 BP Koirala Institute of Health Sciences, Nepal

NATIONAL SCIENTIFIC COMMITTEE

- Dr. Anju Ade (Associate Professor)
 Navodaya Medical College, Raichur, Karnataka
- Dr. E. Venkata Rao (Associate Professor) Community Medicine, Institute of Medical Sciences & SUM Hospital, Bhubaneswar, Orissa.
- Dr. Amit K. Singh (Associate Professor) Community Medicine, VCSG Govt. Medical College, Srinagar – Garhwal, Uttarakhand
- Dr. R G Viveki (Professor & Head) Community Medicine, Belgaum Institute of Medical Sciences, Belgaum, Karnataka
- Dr. Santosh Kumar Mulage (Assistant Professor)
 Anatomy, Raichur Institute of Medical Sciences Raichur(RIMS), Karnataka
- Dr. Gouri Ku. Padhy (Associate Professor) Community and Family Medicine, All India Institute of Medical Sciences, Raipur
- Dr. Ritu Goyal (Associate Professor)
 Anaesthesia, Sarswathi Institute of Medical Sciences, Panchsheel Nagar
- Dr. Anand Kalaskar (Associate Professor)
 Microbiology, Prathima Institute of Medical Sciences, AP
- Dr. Md. Amirul Hassan (Associate Professor)
 Community Medicine, Government Medical College, Ambedkar Nagar, UP
- 10. Dr. N. Girish (Associate Professor) Microbiology, VIMS&RC, Bangalore
- 11. Dr. BR Hungund (Associate Professor) Pathology, JNMC, Belgaum.
- Dr. Sartaj Ahmad (Assistant Professor), Medical Sociology, Department of Community Medicine, Swami Vivekananda Subharti University, Meerut, Uttar Pradesh, India
- Dr Sumeeta Soni (Associate Professor)
 Microbiology Department, B.J. Medical College, Ahmedabad, Gujarat, India

NATIONAL EDITORIAL ADVISORY BOARD

- Prof. Sushanta Kumar Mishra (Community Medicine) GSL Medical College – Rajahmundry, Karnataka
- Prof. D.K. Srivastava (Medical Biochemistry)
 Jamia Hamdard Medical College, New Delhi
- Prof. M Sriharibabu (General Medicine) GSL Medical College, Rajahmundry, Andhra Pradesh
- Prof. Pankaj Datta (Principal & Prosthodentist) Indraprastha Dental College, Ghaziabad

NATIONAL EDITORIAL ADVISORY BOARD

- Prof. Samarendra Mahapatro (Pediatrician)
 Hi-Tech Medical College, Bhubaneswar, Orissa
- Dr. Abhiruchi Galhotra (Additional Professor) Community and Family Medicine, All India Institute of Medical Sciences, Raipur
- Prof. Deepti Pruthvi (Pathologist) SS Institute of Medical Sciences & Research Center. Davangere. Karnataka
- 8. Prof. G S Meena (Director Professor)
 Maulana Azad Medical College, New Delhi
- Prof. Pradeep Khanna (Community Medicine)
 Post Graduate Institute of Medical Sciences, Rohtak, Haryana
- Dr. Sunil Mehra (Paediatrician & Executive Director) MAMTA Health Institute of Mother & Child, New Delhi
- Dr Shailendra Handu, Associate Professor, Phrma, DM (Pharma, PGI Chandigarh)
- Dr. A.C. Dhariwal: Directorate of National Vector Borne Disease Control Programme, Dte. DGHS, Ministry of Health Services, Govt. of India, Delhi

Print-ISSN: 0976-0245-Electronic-ISSN: 0976-5506, Frequency: Quarterly (Four issues per volume)

Indian Journal of Public Health Research & Development is a double blind peer reviewed international journal. It deals with all aspects of Public Health including Community Medicine, Public Health, Epidemiology, Occupational Health, Environmental Hazards, Clinical Research, and Public Health Laws and covers all medical specialties concerned with research and development for the masses. The journal strongly encourages reports of research carried out within Indian continent and South East Asia.

The journal has been assigned International Standards Serial Number (ISSN) and is indexed with Index Copernicus (Poland). It is also brought to notice that the journal is being covered by many international databases. The journal is covered by EBSCO (USA), Embase, EMCare & Scopus database. The journal is now part of DST, CSIR, and UGC consortia.

Website: www.ijphrd.com

©All right reserved. The views and opinions expressed are of the authors and not of the Indian Journal of Public Health Research & Development. The journal does not guarantee directly or indirectly the quality or efcacy of any product or service featured in the advertisement in the journal, which are purely commercial.

Editor

Dr. R.K. Sharma

Institute of Medico-legal Publications Logix Office Tower, Unit No. 1704, Logix City Centre Mall, Sector- 32, Noida - 201 301 (Uttar Pradesh)

Printed, published and owned by

Dr. R.K. Sharma

Institute of Medico-legal Publications Logix Office Tower, Unit No. 1704, Logix City Centre Mall, Sector- 32, Noida - 201 301 (Uttar Pradesh)

Published at

Institute of Medico-legal Publications

Logix Office Tower, Unit No. 1704, Logix City Centre Mall, Sector- 32, Noida - 201 301 (Uttar Pradesh)



Indian Journal of Public Health Research & Development

www.ijphrd.com

Contents
Volume 10, Number 3 March 2019
A Study of Financial Counselling and Estimation of Variance between Estimated Bill and Actual Bill of Cardiac Cath Lab Brig. A P Pandit, Harshada Tambe
2. Alienation, Attachment Style, and Alcohol Addiction "A Study of Young Women Habitual Drinkers" Bhawani Singh Rathore, Uma Joshi
3. Challenging Issues in Health Economics
4. Examining the Impact of Development Projects' on the Project affected People: A Study in Kalinganagar Industrial Estate, Odisha, India
5. Effectiveness of Hand Muscle Strenthening Exercises on Dominant Hand Pinch Strength in Smartphone Addicted Collegiate
6. Effect of Traction Versus Counter Traction with Conservative Treatment on Adhesive Capsulitis 3 Divya Nitin Lad, S. Anandh
7. Effficacy of Probiotic Drink Containing Lactobacillus Casei Shirota Strain on Factors affecting Dental Caries
8. Health Expectancy Under Dynamic Set Up for India and its Selected States
9. IoT based E-Critical Care Unit for Patients In-Transit
10. The Influence of Moderate Intensity Exercises on Fatigue and Quality of Life in Cancer Patients Undergoing Chemotherapy
11. The Effectiveness of Planned Educational Intervention on Knowledge and Skills in Home Care Manageme among Care Givers of Patients with Spinal Cord Injury
12. Microbiological Evaluation of Polytetrafluoroethylene (PTFE) Tape, Cellulose Sponge and Cotton as Spacer Materials Combined with Intracanal Medicament- An in Vitro Study
13 Socio-Demographic and Cultural Factors Influencing Treatment Outcomes among Patients with

169.	Mapping of Tuberculosis (TB) Prevalence in Padang City	877
170.	History of Children and Malnutrition Status in Magetan	882
171.	Mother's Knowledge, Attitude, and Practice of Exclusive Breastfeeding	887
172.	Factors Associated with Cerumen Impaction in the Coastal Elementary Schools (Case Study in 1st Grade of five Elementary Schools, Bandarharjo Health Center's Work Area, in North Semarang)	893
173.	Exploring the Compliance Test for X-ray in Health Facilities Security of Makassar Region	899
174.	A Comparative Analysis between Integrated Occupational Safety and Health Management System in a Support Mining Company and the Indonesian Mining Safety Management System Effendi Herman Ratar, Hanifa M. Denny, Mohammad Zen Rahfiludin	904
175.	The Correlation between Regulation Understanding by Inter-Professional first 1000 days of Life Health Workers and the Acceleration of Toddler Stunting Prevention	911
176.	Cadres Outreach Program to Mothers Improve Nutritional Status of Under Two Year Old Children Martha Irene Kartasurya, Dina Rahayuning Pangestuti, Aditya Kusumawati, Resa Ana Dina, Julia Pertiwi, Doddy Izwardy, Anung Sugihantono	917
177.	Positive Emotion, Engagement and Meaning of Life of the Elderly in <i>Pesantren</i> (Islamic Boarding School) <i>Dinie R. Desiningrum, Yeniar Indriana, Suparno</i>	923
178.	Health Professional's Perception toward Impact of Hospital Accreditation on Quality of Care in Asia: A Systematic Review	929
179.	The Evaluation of Acute Appendicitis Clinical Pathway	935
180.	Determinants of the Pornography Exposure Effects on Junior and Senior High School Adolescence in Sanggau District, West Kalimantan	941
181.	Psychoeducation <i>Dzikr</i> Reduces Perceived Stress and Postpartum Depression Syndromes on Primiparous Women	946
182.	Reduction Sugar of Tuber Paste Flour Additional α-Amylase from <i>Lc. mesenteroides</i> EN17-11 and <i>Fr. fructosus</i> EN17-20 to Protect People from Diabetes Mellitus	952
183.	The Impact of Counseling on the Improvement of Nutritional Knowledge and Physical Activities on Women Prisoners (A Study at Women Penitentiary Institution Class II A Semarang)	958

Exploring the Compliance Test for X-ray in Health Facilities Security of Makassar Region

Alan Tanjung Aji Prastowo¹, Yuliani Setyaningsih², Hanifa M. Denny²

¹Master of Health Promotion, Faculty of Public Health, ²Faculty of Public Health, Diponegoro University

ABSTRACT

The initial observation on compliance test for X-ray facilities reported that some of them were failed to meet the obligation of compliance test provided by the Health Facilities Security of Makassar Region or BPFK Makassar. The purpose of this study was to explore the reasons for their failure to fulfill the obligation on X-ray facilities compliance and services. This research employed qualitative methods. The ten informants of 9 primary informants and 1 secondary informant from the X-ray health facilities in Makassar of South Sulawesi Island were interviewed. The result behind the success in their compliance test was mainly due to the presence of the supervision team from the Nuclear Energy Regulatory Agency of Indonesia. Additionally, another factor such as the presence of internal and external supervision also played significant support to be able to fulfill the compliance test for X-ray facilities. At the other side, inadequate funding and lack of understanding about the purpose and obligation of the compliance test were mostly informed by the respondents as the main reasons for their failure to comply with the calibration test of the X-ray facilities. Therefore, external and internal supervision should be strengthened to increase the compliance test among X-ray health facilities in Makassar.

Keywords: Compliance Test, X-ray, Health Facilities, Nuclear Energy Regulatory Agency

INTRODUCTION

The Minister of Health of Indonesia Regulation Number. 432/MENKES/SK/IV/2007 on occupational safety and health management for the hospital stipulated that radiation is one of potential physical hazards to be concerned by planning, organizing, implementing, and controlling aiming to cultivate occupational safety and health in the hospital especially in radiology examination.

The ionizing radiation that involves the human body can cause deterministic effects such as skin erythema, cataract, sterility, nausea, diarrhea, fetal death, stochastic effects such as cancer, and hereditary defects. The radiation protection is useful to prevent the deterministic effects and to decrease the probability of stochastic effects among workers.

Corresponding Author:

Alan Tanjung Aji Prastowo Master of Health Promotion, Faculty of P ublic Health, Diponegoro University

Phone: +6285642162727

Email: alantanjungajiprastowo@gmail.com

Meanwhile, the compliance test is an assurance of the X-ray used to prevent unnecessary radiation dose on the patients, the workers and the person who may around. The x-ray machine reliability assurance in Indonesia is regulated by BAPETEN Regulation Number 9/2011 regarding the implementation of the compliance test and the Indonesian Minister of Health Regulation Number 54/2015 on the medical equipment compliance test. According to BAPETEN Regulation Number 9/2011, every health facilities that request a new permit or an extension of the license of X-ray machine is mandated to apply for the compliance test. If a case occurs.²

The services of the compliance test applied by BPFK Makassar include 10 zones in eastern Indonesia: South Sulawesi, West Sulawesi, Central Sulawesi, Southeast Sulawesi, Gorontalo, North Sulawesi, North Maluku, Maluku, Papua, and West Papua. From 2015 to 2017 some 124, 114, 134, and 145 facilities around the Eastern part of Indonesian territory had the compliance test once or twice. The Director General of Health Services of the Ministry of Health Republic Indonesia showed that 315 hospitals under BPFK Makassar's coverage of General Hospitals and Specialty Hospitals had the X-ray

facilities.³ Unfortunately, many of them did not meet the standard of twice compliance tests annually. According to a theory, the input, the process, and the output in health administration is unique according to who and how to deal with their limitation.⁴

There are some aspects such as health facilities, fund, policy, and feedback that are in need to boost the fulfillment of twice compliance tests for the X-ray facilities annually.

METHOD

This research was qualitative research. The government's X-ray facilities that implement twice tests continuously or called 'SP.A,' while the private X-ray facilities are called 'SP.B.' The government health facilities that do not implement the twice tests

periodically are called 'SP.C,' and the private ones are called 'SP.D.' The informants consisted of management staff and a technical worker from both sides, the X-ray facilities and the BPFK of Makassar. An in-depth interview was employed to gather data. A total of ten informants from the X-ray health facilities in Makassar of South Sulawesi Island were interviewed, and their responses were re-check using triangulation with the BPFK staff and technical operator.

RESULTS

Data Analysis of BPFK Makassar: According to BPFK Makassar services, the compliance test from 2014 to 2017 increased in term of the number of test participants, besides the decreasing number of health facilities from 124 facilities to 114 from 2014 to 2015. The result of the implementation of the test can be shown in Table 1..

Table 1: A number of health facilities implement the compliance test and calibration in the year of 2014–2018

Province	2014	2015	2016	2017	2018*
North Sulawesi	8	8	8	10	6
South East Sulawesi	9	5	7	12	2
Central Sulawesi	11	7	11	15	8
South Sulawesi	57	51	66	55	17
West Sulawesi	4	4	3	7	0
West Papua	6	1	3	2	1
Papua	7	9	7	9	2
North Maluku	5	4	7	4	4
Maluku	6	2	5	8	2
Gorontalo	6	14	6	9	2
Outside The Working Area of BPFK Makassar	5	9	11	14	4
Total	124	114	134	145	48

^{*}The data until 2018 May

South Sulawesi is a province with the largest number of health facilities in East Indonesia. Almost half or 44.29% implemented the compliance test with BPFK Makassar annually. In 2017, there was an increasing number of compliance test participants up to 70% of X-ray facilities.

Table 2: The number of Health Facilities based on both possession of government and private from year of 2014-2018

Health Facilities	2014	2015	2016	2017	2018*
Central Government	4	4	6	5	4
Province Government	12	11	12	17	3
County/City Government	48	40	56	57	16
Another Minister	2	2	2	3	0
Private Hospitals	28	24	29	30	7
Clinics	24	26	14	24	7
BUMN	1	2	3	0	0
Companies	2	2	4	2	4
TNI/POLRI	3	3	8	7	7
Total	124	114	134	145	48

^{*}The data until 2018 May

In the input, process, and output are shown in figure 1. the X-ray facilities attached to the government showed an increase in compliance test fulfillment, but the private ones didn't show any progress.




Figure 1: Scheme of the application of the compliance test and calibration with Systems Theory

Interviews Result

Input: All management and staff in 'SP.A,' SP.B,' and SP.C' knew the latest regulation of the compliance test that should be fulfilled biennial or once every two years along with the extension of the license. Funding was the main reason for failure in the fulfillment of the compliance test.

Subsequently, the standard operating procedure had to be implemented for the X-ray facilities if the facility sought the compliance test and licensed to operate it.

The 'SP.A,' 'SP.B' and 'SP.D' obtained the compliance test with BPFK Makassar. However, 'SP.C' had calibrated with private labs in 2017. Financially, one of the respondents said that the cost of medical equipment calibration by BPFK Makassar is too expensive.

Process: Planning for the test was conducted annually. The X-ray Facilities with ('SP.A') could not schedule the test because they had no staff to be in charge of the implementation achievement. The 'SP.B' code of X-ray facility usually set a reminder on the 90th days before the expiration of the X-ray license.

Staffing had been applied by the management 'SP.A,' 'SP.B,' and 'SP.C' by employing staff for coordinating this implementation of the compliance test and calibration. A radiation protection officer directly involved in the

process of the test. The 'SP.D' code for the radiology clinics own by private did not have the coordinator for their compliance test. Therefore, a radiation protection officer should take in charge of compiling the documents of the standard of the permission for the compliance test.

Nuclear Energy Regulatory Agency (BAPETEN) once conducted inspections in all of the research subject. The compliance test related to the accreditation requirements as well as a mean to protect the patient. Subsequently, marketing aspects are also taking advantages of the compliance aspects of the X-ray facilities.

Output: Research subjects 'SP.A' and 'SP.B' applied the compliance test and calibration periodically, due to human resources awareness on the rules and the benefit of the compliance test. Currently, a significant fund is available to cover the test by BPFK Makassar. However, the calibration was not implemented in 'SP.C' due to the insignificant fund, even though 'SP.C' was supported by the input system.

The implementation of the calibration periodically could not be applied in 'SP.D' due to the lack of understanding the implementation of calibration for every year, but mere applying the compliance test for the standard of the permission for two year.

DISCUSSIONS

Health facilities in two areas such as both West and South Sulawesi could apply the compliance test and calibration due to the close distance with BPFK Makassar. However, health facilities in West Papua could use this test least due to this reason. West Papua is a large area, where most of the regions are wildlife jungle. Technical Unit in Papua could not afford this compliance test, so Papua and West Papua must invite the BPFK testers. According to Weeren (2016), convenience places affected partially.⁵

Input: The supervision from Nuclear Energy Regulatory Agency towards the X-ray facilities and the requirement for a license are the contributor for their implementation of the compliance test. A sanction is enforced according to the Regulation on nuclear energy Number 10/1997.⁶

The compliance test and calibration of the X-ray facility are related to the availability of funding as informed by the respondent. Lack of financing among 'SP.C' X-ray facilities was the reason for not being able

to fulfill the compliance test. Although the commitment of the organization presence, without sufficient funding, the compliance test was trying to enforce. Government Regulation Number 21/2013 stipulated that funding should cover the accommodation, transportation, and daily allowance for the person who conducted the test. However, many X-ray facilities perceived a significant burden to accommodate the compliance test cost. 8

The BPFK Makassar has the dimension of service quality with top categories and the aspect of services quality with sufficient excellent facilities that put them in a pressure to work harder in convincing and advocating X-ray facilities to obtain the compliance test.

Process: The correct information system can strengthen the importance of the planning, and ensure the preparation run well. SPB applies the information system through plan ordered in the last year, concerning the limit time of the certificate, so SPB can afford to employ the calibration test punctually.

As BAPETEN Chairman's Regulation No. 4/2013, health facilities utilizing the source of ionizing radiation should have the radiation protection officers. Health facilities had the radiation protection officers even though the officers were not the permanent employees. The test coordinator pointed is to be expected to ease the implementation of the compliance test and calibration periodically. Even though SPC had the test coordinators but the fact of the lack of fund caused not the application of the compliance test and calibration regularly. The role of radiation protection officer is essential, but if their obligations are violated, merely used as the standard of nuclear energy permission, so the sanction will be conducted as mentioned in Regulation Number 10/1997.

The most effective supervision is an inspection from BAPETEN. However, it is just limited to the implementation in the compliance test for the permission. People should concern labeling applied by BAPETEN for their common safety. In the calibration implementation, the factor of accreditation assessments done by health facilities becomes the supporting factor of the execution of the compliance test and calibration routinely. SPB also did the internal supervision through the internal audit for every year. This audit is a management step as the supervision effort if disobedience is found in the result of the internal review so that the top management can stop the operational examination permission. The

health facilities supervising the implementation of compliance test and calibration to reduce doses received by the officer, although the addition of radiation doses is also influenced by use inappropriate personal protective equipment.¹¹

Output: The implementation of the compliance test and calibration routinely applied by 'SP.A' and 'SP.B' was affected by the changed input: knowledge and obedience of the human resources towards the regulation of these two tests, the availability of sufficient fund, of procedure, run well, changed the process of planning, staffing, and supervising can increase this implementation.

The application of radiation safety management by applying the quality control of x-ray machine in health facilities should be conducted as it mentioned in its regulations.¹²

CONCLUSIONS

The X-ray facilities without complying with the regulation on compliance test existed in the coverage area of BPFK Makassar.

The inspection by BAPETEN supported the implementation of the compliance test. The requirement to extend the license operation of the X-ray facility contributed to comply with the compliance test.

Source of Funding: Funding also played an essential role in meeting the compliance test mandatory.

Conflict of Interest: The Authors declare no conflict of interest in this article. This research is self funded.

Ethical Clearance: The Ethical Clearance was obtained from the Committee of Ethical Research No. 069/EC/FKM/2018 on 04 June 2018

REFERENCES

- 1. Goodman TR. Ionizing Radiation Effects and Their Risk to Humans. Am Coll Radiol. 2010;
- 2. Oey, Gardiner M. Study Of The Role Of the Indonesian Institute Indonesia (LIPI) in Brdging between research and development policy. 2011.
- 3. Handayani PW, Hidayanto AN, Pinem AA, Hapsari IC, Sandhyaduhita PI, Budi I. Acceptance model of a Hospital Information System. Int J Med Inform. 2017;

- 4. Novick, Morrows. Public Health Administration: Principles for Population-Based Management. USA: Michael Brown; 2013. 53-78 p.
- Weeren RJ van, Williams WL, Hardcastle MJ, Shimwell TW, Rafferty DA, Sabater J, et al. LOFAR FACET CALIBRATION. Astrophys J Suppl Ser. 2016;
- Arifin Z. Challenges to Implementation of Knowledge Management on Drafting of Nuclear Regulation Indonesia. 2007;
- Handayani PW, Hidayanto AN, Sandhyaduhita PI, Kasiyah, Ayuningtyas D. Strategic hospital services quality analysis in Indonesia. Expert Syst Appl. 2015;
- 8. State S. Law of the Republic of Indonesia. 2002;(19):1–19. Available from: http://portal.unesco.org/culture/en/

- files/30382/11424187703id_copyright_2002_en.pdf/id copyright 2002 en.pdf
- 9. Raghunathan T., W.R.King. The Impact Of Information Systems Planning on The Organization. Omega 16.2; 1988. 85-93 p.
- 10. Utami S., Putero S. Regulatory Framework for the Decommisioning of Indonesian Nuclear Facilities. J Nucl Eng Radiat Sci. 2017;3(4).
- 11. Mohamed A, Elamin T. Radiation Safety Awareness and Practice in Sudanese Medical Facilities: A Descriptive. Int J Sci Res ISSN (Online Index Copernicus Value Impact Factor. 2013;
- 12. Briggs-Kamara M, Okoye P. Radiation safety awareness among patients and radiographers in three hospitals in Port Harcourt. Am J Sci Ind Res. 2013;